

# Certificate

The Certification Body of  
**TÜV Rheinland LGA Products GmbH**

hereby certifies that the organization  
**Anhui Tiankang Medical  
Technology Co., Ltd.**  
**No. 228 Weiye Road**  
**Economic Development Zone**  
**Tianchang City**  
**239300 Anhui**  
**China**

has established and applies a quality management system for medical devices  
for the following scope:

**Manufacture and Distribution of Medical Devices**

**(see attachment for products and additional site included)**

Proof has been furnished that the requirements specified in

**EN ISO 13485:2016**

are fulfilled. The quality management system is subject to yearly surveillance.

Effective Date: 2018-09-17  
Certificate Registration No.: SX 60131062 0001  
An audit was performed. Report No.: 15096003 004  
This Certificate is valid until: 2021-09-16

Certification Body



Date 2018-09-17



**TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg**

Tel.: +49 221 806-1371 Fax: +49 221 806-3935 e-mail: cert-validity@de.tuv.com <http://www.tuv.com/safety>

**1.0 EC Declaration of Conformity**

Name: Anhui Tiankang Medical Technology Co., Ltd.

Manufacturer: Add: No. 228 Weiyi Road, Economic Development Zone, Tianchang City, 239300 Anhui, China

European Name: MedPath GmbH

Representative: Add: Mies-van-der-Rohe-Strasse 8, 80807 Munich, Germany

Product Name: Single-use Medical Face Mask  
(Type I, Type II, Type IIR)

Object of the Types/Sizes declaration: With earloop or with earloop(flat earloop) or with earloop(folded earloop) or with tie coverall: 12.5cm x 9cm, 14.5cm x 9cm, 16.5cm x 9cm, 17cm x 8cm, 17cm x 9cm, 17cm x 9.5cm, 17cm x 10cm, 17.5cm x 8cm, 17.5cm x 9cm, 17.5cm x 9.5cm, 17.5cm x 10cm, 18cm x 8cm, 18cm x 9cm, 18cm x 9.5cm, 18cm x 10cm, 19.5cm x 8cm, 19.5cm x 9cm, 19.5cm x 9.5cm, 19.5cm x 10cm

UMDNS Code: 12-447

Classification (MDD, Annex IX): I, rule 1

Conformity Assessment Route: Annex VII

We herewith declare in sole responsibility that the above mentioned products meet the transposition into national law, the provisions of the following EC Council Directives and Standards such as EN 14683:2019+AC:2019 etc. All supporting documentations are retained under the premises of the manufacturer. We are exclusively responsible for the declaration of conformity.

**DIRECTIVES**

General applicable directives:

Medical Device Directive: COUNCIL DIRECTIVE 93/42/EEC

Start of CE Marking: 2020-03-25

Place of Issue: Tianchang, CHINA


Date of Issue: 2020-04-23

Signature:



Mr. Baodong Bai

Position: General Manager





MedPath

# EC-Registration Certificate

Directive 93/42/EEC on Medical Devices (MDD), Article 14

No. R A001 29 Rev. 01

Manufacturer: Anhui Tiankang Medical Technology Co., Ltd.

No. 228 Weiyi Road, Economic Development Zone,  
Tianchang City, 239300 Anhui, China

Product

See Appendix A

Category(ies):



This is to certify that, in accordance of the Medical Device Directive 93/42/EEC (amended by 2007/47/EC), MedPath GmbH agrees to perform all duties and responsibilities as the Authorized Representative for the aforementioned manufacturer as stipulated and demanded by the aforementioned Directive. The German Competent Authority is notified of the manufacturer's medical device(s) and has allocated registration numbers shown in Appendix A. The manufacturer has provided MedPath GmbH with the appropriate Declaration(s) of Conformity confirming that the medical device(s) fulfills/fulfill the applicable requirements of the aforementioned Directive.

**MedPath GmbH**

Mies-van-der-Rohe-Strasse 8 · D-80807 München  
Tel. 089-189174474 · Fax 089-54858884



Date, 2020-03-28

MedPath GmbH



MedPath

## Appendix A: Product Category(ies)

No.	Name	Class	UMDNS Code	Form No.	Registration No.
1	Single-use Medical Face Mask	I	12-447	00297923	to be issued
2	Non-woven Coveralls	I	15-223	00297925	to be issued
3	Non-woven Isolation Gowns	I	15-037	00297927	to be issued



**MedPath GmbH**  
Mies-van-der-Rohe-Strasse 8 • D-80807 München  
Tel. 089-189174474 • Fax 089-54858884



# **Allgemeine Anzeigepflicht nach §§ 25 und 30 Abs. 2 MPG** **General Obligation to Notify pursuant to §§ 25 and 30 (2) Medical Devices Act, MPG**

## **Formblatt für Medizinprodukte, außer In-vitro-Diagnostika** **Form for Medical Devices except In Vitro Diagnostic Medical Devices**

<b>Zuständige Behörde / Competent authority</b>	
Code <b>DE/CA61</b>	
Bezeichnung / Name <b>Regierung von Oberbayern</b>	
Staat / State <b>Deutschland</b>	Land / Federal state <b>Bayern</b>
Ort / City <b>München</b>	Postleitzahl / Postal code <b>80534</b>
Straße, Haus-Nr. / Street, house no. <b>Maximilianstraße 39</b>	
Telefon / Phone <b>+49-89-21760</b>	Telefax / Fax <b>+49-89-21762914</b>
E-Mail / E-mail <b>medizinprodukteanzeigeverfahren@reg-ob.bayern.de</b>	

<b>Anzeige / Notification</b>	
Registrierdatum bei der zuständigen Behörde Registration date at competent authority	Registriernummer / Registration number
Typ der Anzeige / Notification type <input checked="" type="checkbox"/> Erstanzeige / Initial notification <input type="checkbox"/> Änderungsanzeige / Notification of change <input type="checkbox"/> Widerrufsanzeige / Notification of withdrawal	
Frühere Registriernummer bei Änderungs- und Widerrufsanzeige Previous registration number if notification has been changed or withdrawn	
Anzeigender nach § 25 MPG / Reporter pursuant to § 25 Medical Devices Act, MPG <input type="checkbox"/> Hersteller / Manufacturer <input checked="" type="checkbox"/> Bevollmächtigter / Authorised Representative <input type="checkbox"/> Einführer / Importer <input type="checkbox"/> Verantwortlicher für das Zusammensetzen von Systemen oder Behandlungseinheiten nach § 10 Abs. 1 und 2 MPG \ Assembler of systems or procedure packs pursuant to § 10 (1) and (2) Medical Devices Act, MPG <input type="checkbox"/> Betrieb oder Einrichtung (aufbereiten) nach § 25 Abs. 1 MPG i. V. m. § 4 Abs. 2 MPBetreibV Institution (processing) pursuant to § 25 (1) Medical Devices Act, MPG in connection with § 4 (2) MPBetreibV <input type="checkbox"/> Betrieb oder Einrichtung (sterilisieren) nach § 25 Abs. 2 i. V. m. § 10 Abs. 3 MPG Institution (sterilizing) pursuant to § 25 (2) in connection with § 10 (3) Medical Devices Act, MPG	

<b>Anzeigender / Reporting organisation (person)</b>	
Code	<b>DE/0000047823</b>
Bezeichnung / Name	<b>MedPath GmbH</b>
Staat / State	<b>Deutschland</b>
Land / Federal state	<b>Bayern</b>
Ort / City	<b>München</b>
Postleitzahl / Postal code	<b>80807</b>
Straße, Haus-Nr. / Street, house no. <b>Mies-van-der-Rohe-Strasse 8</b>	
Telefon / Phone	<b>089 189174474</b>
Telefax / Fax	
E-Mail / E-mail <b>info@medpath.pro</b>	

<b>Hersteller / Manufacturer</b>	
Bezeichnung / Name	<b>Anhui Tiangkang Medical Technology Co., Ltd.</b>
Staat / State	<b>CN</b>
Ort / City	<b>Tianchang City</b>
Postleitzahl / Postal code	<b>239300</b>
Straße, Haus-Nr. / Street, house no. <b>No. 228 Weiye Road, Economic Development Zone</b>	
Telefon / Phone	<b>+86 13705505106</b>
Telefax / Fax	
E-Mail / E-mail <b>zy@tkmedical.com</b>	

<b>Sicherheitsbeauftragter für Medizinprodukte nach § 30 Abs. 2 MPG 9)</b> <b>Safety officer for medical devices pursuant to § 30 (2) Medical Devices Act, MPG</b>	
Bezeichnung / Name	<b>Zheng Mei c/o MedPath GmbH</b>
Staat / State	<b>Deutschland</b>
Land / Federal state	<b>Bayern</b>
Ort / City	<b>München</b>
Postleitzahl / Postal code	<b>80807</b>
Straße, Haus-Nr. / Street, house no. <b>Mies-van-der-Rohe-Strasse 8</b>	
Telefon / Phone	<b>089 189174474</b>
Telefax / Fax	<b>089 5485 8884</b>
E-Mail / E-mail <b>info@medpath.pro</b>	

<b>Vertreter / Deputy (optional)</b>	
	Bezeichnung / Name
Telefon / Phone	Telefax / Fax
E-Mail / E-mail	
<input checked="" type="checkbox"/> Erstanzeige / Initial notification <input type="checkbox"/> Änderungsanzeige / Notification of change	



Medizinprodukt (Erstmaliges Inverkehrbringen) / Medical device (First placing on the market)	
Klasse / Class	<input checked="" type="checkbox"/> I <input type="checkbox"/> I - steril / sterile <input type="checkbox"/> I - mit Messfunktion / with measuring function <input type="checkbox"/> I - steril und mit Messfunktion / sterile and with measuring function <input type="checkbox"/> IIa <input type="checkbox"/> IIb <input type="checkbox"/> III <input type="checkbox"/> III - hergestellt unter Verwendung von Gewebe tierischen Ursprungs im Sinne der Verordnung (EU) Nr. 722/2012 manufactured utilising tissues of animal origin in terms of Commission Regulation (EU) No 722/2012 <input type="checkbox"/> Aktives implantierbares Medizinprodukt / Active implantable medical device <input type="checkbox"/> Aktives implantierbares Medizinprodukt - hergestellt unter Verwendung von Gewebe tierischen Ursprungs im Sinne der Verordnung (EU) Nr. 722/2012 Active implantable medical device - manufactured utilising tissues of animal origin in terms of Commission Regulation (EU) No 722/2012
App (Software auf mobilen Endgeräten)	<input type="checkbox"/> ja / yes <input checked="" type="checkbox"/> nein / no
Nummer(n) der Bescheinigung(en) / Certificate number(s)	
Handelsname des Produktes / Trade name of the device	<b>Single-use Medical Face Mask</b>
Produktbezeichnung / Name of device	
Nomenklaturcode / Nomenclature code	<b>12-447</b>
Nomenklaturbezeichnung / Nomenclature term	<b>Maske</b>
Kategoriecode / Category code	<b>10</b>
Kategorie / Category	<b>Produkte zum Einmalgebrauch</b>
Kurzbeschreibung deutsch / German short description	
Kurzbeschreibung englisch / English short description	



<b>Medizinprodukte (Aufbereiten) / Medical devices (Reprocessing)</b>	
	<input type="checkbox"/> Semikritische Medizinprodukte / Semicritical medical devices <input type="checkbox"/> Gruppe A / Group A <input type="checkbox"/> Gruppe B / Group B
	<input type="checkbox"/> Kritische Medizinprodukte / Critical medical devices <input type="checkbox"/> Gruppe A / Group A <input type="checkbox"/> Gruppe B / Group B <input type="checkbox"/> Gruppe C / Group C Nummer der Bescheinigung / Certificate number
	Sterilisationsverfahren / Sterilisation procedures <input type="checkbox"/> Dampfsterilisation / Steam sterilisation <input type="checkbox"/> Gassterilisation / Gas sterilisation <input type="checkbox"/> Strahlensterilisation / Radiation sterilisation <input type="checkbox"/> andere / others Angewandtes Verfahren / Applied procedure

Ich versichere, dass die Angaben nach bestem Wissen und Gewissen gemacht wurden.  
I affirm that the information given above is correct to the best of my knowledge.

Ort City	<b>München</b>	Datum Date	<b>2020-03-28</b>
		Name	<b>Song Wang</b>
			Unterschrift Signature

<b>Bearbeitungsvermerke / Processing notes</b> Nur von der zuständigen Behörde auszufüllen / To be filled in only by the competent authority	
Bearbeiter / Person responsible	Telefon / Phone

**SUBJECT** Physical & Microbiological Test

**TEST LOCATION** TÜV SÜD China

TÜV SÜD Products Testing (Shanghai) Co., Ltd.  
B-3/4, No.1999 Du Hui Road, Minhang District  
Shanghai 201108, P.R. China

**CLIENT NAME** Anhui Tiankang Medical Technology Co.,Ltd

**CLIENT ADDRESS** No.228 Wei yi Road Economic Development Zone Tianchang City 239300  
Anhui China

**TEST PERIOD** 01-May-2020~12-May-2020

**Prepared By**

Bella Xu

(Bella Xu)  
Report Drafter

**Authorized By**



(Leo Liu)  
Authorized Signatory

**Note:** (1) General Terms & Conditions as mentioned overleaf. (2) The results relate only to the items tested. (3) The test report shall not be reproduced except in full without the written approval of the laboratory. (4) Without the agreement of the laboratory, the client is not authorized to use the test results for unapproved propaganda.

**Chemical/Microbiology Laboratory:**  
TÜV SÜD Products Testing (Shanghai) Co., Ltd.  
B-3/4, No.1999 Du Hui Road, Minhang District  
Shanghai  
201108  
P.R. China

Phone : +86 (21) 6037 6375  
Fax : +86 (21) 6037 6345  
Email: food.chem@tuv-sud.cn  
Webpage: www.tuv-sud.cn

**Regional Head Office:**  
TÜV SÜD Certification and Testing  
(China) Co., Ltd.  
No.151 Heng Tong Road Shanghai  
200 070 P.R.China

**TUV**<sup>®</sup>

## TEST REPORT

Sample Description : Single-use Medical Face Mask  
Sample Quantity : 50 pieces  
Lot Number/Batch Code : 200416  
Specification : With earloop  
Size : /  
Brand Name : /

Remark: The above information was provided by applicant.

### Summary of Test Results

No.	Test Item	Test Method	Test Standard Type I	Judgement
1	Bacterial Filtration Efficiency Test (BFE), %	EN 14683:2019+AC:2019(E) Annex B	$\geq 95$	Pass
2	Differential Pressure Test (Pa/cm <sup>2</sup> )	EN 14683:2019+AC:2019(E) Annex C	$< 40$	Pass
3	Microbial Cleanliness Test (CFU/g)	EN 14683:2019+AC:2019(E) Annex D	$\leq 30$	Pass

Note: Pass = Meet customer requirements;

Fail = Fail customer requirements;

# = No comment;

N.D. = Not detected.

### Photo of Samples



## Results

No.	Test Item	Test Result
1	Bacterial Filtration Efficiency (BFE) Test	Specimen 1#: 99.7% Specimen 2#: 99.6% Specimen 3#: 99.7% Specimen 4#: 99.7% Specimen 5#: 99.6%
2	Differential Pressure Test	27.3 Pa/cm <sup>2</sup>
3	Microbial Cleanliness Test	Specimen 1#: 6 CFU/g Specimen 2#: 6 CFU/g Specimen 3#: 4 CFU/g Specimen 4#: 3 CFU/g Specimen 5#: 3 CFU/g

### Bacterial Filtration Efficiency (BFE) Test

#### 1. Purpose

For evaluating the bacterial filtration efficiency (BFE) of mask.

#### 2. Sample description was given by client

Sample description : Single-use Medical Face Mask  
Specification : With earloop  
Lot Number : 200416  
Sample Receiving Date : 2020-05-01

#### 3. Test Method

EN 14683:2019+AC:2019(E) Annex B

#### 4. Apparatus and materials

- 4.1 *Staphylococcus aureus* ATCC 6538.
- 4.2 Peptone water.
- 4.3 Tryptic Soy Broth(TSB).
- 4.4 Tryptic Soy Agar(TSA).
- 4.5 Bacterial filtration efficiency test apparatus.
- 4.6 Six-stage viable particle Anderson sampler.
- 4.7 Flow meters.

#### 5. Test specimen

- 5.1 As requested by client, take a total of 5 test specimens.
- 5.2 Prior to testing, condition all test specimens for a minimum of 4 h at (21±5)°C and (85±5)% relative humidity.

## 6. Procedure

- 6.1 Preparation of the bacterial challenge: Dilute the culture in peptone water to achieve a concentration of approximately  $5 \times 10^5$  CFU/mL.
- 6.2 Adjust the flow rate through the Anderson sampler to 28.3 L/min.
- 6.3 Deliver the challenge to the nebulizer using a syringe pump. Purge tubing and nebulizer of air bubbles.
- 6.4 Perform a positive control run without a test specimen to determine the number of viable aerosol particles being generated. The mean particle size (MPS) of the aerosol will also be calculated from the results of these positive control plates.
  - 6.4.1 Initiate the aerosol challenge by turning on the air pressure and pump connected to the nebulizer. Immediately begin sampling the aerosol using the Anderson sampler.
  - 6.4.2 Time the challenge suspension to be delivered to the nebulizer for 1 min.
  - 6.4.3 Time the air pressure and Anderson sampler to run for 2 min.
  - 6.4.4 At the conclusion of the positive control run, remove plates from the Anderson sampler.
- 6.5 Place new agar plates into Anderson sampler and clamp the test specimen into the top of the Anderson sampler, with the inside of the specimen facing towards the bacterial challenge (test area: 77cm<sup>2</sup>).
- 6.6 Repeat the challenge procedure for each test specimen.
- 6.7 Repeat a positive control after completion of the sample set.
- 6.8 Perform a negative control run by collecting a 2 min sample of air from the aerosol chamber. No bacterial challenge should be pumped into the nebulizer during the collection of the negative control.
- 6.9 Incubate agar plates at (37±2)°C for (20 to 52) h.
- 6.10 Count each of the six-stage plates of the Anderson sampler.

## 7. Calculation

Total the count from each of the six plates for the test specimens and positive controls, as specified by the manufacture of Anderson sampler. The filtration efficiency percentages are calculated as follows:

$$BFE = (C - T) / C \times 100$$

*T* is the total plate count for the test specimen.

*C* is the mean of the total plate counts for the two positive controls.

## 8. Test results\*

<b>P Value</b> <b>Stage</b> <b>Number</b>	<b>Positive</b> <b>Control (A)</b>	<b>Positive</b> <b>Control (B)</b>	<b>Negative</b> <b>Control</b>	<b>Specimen</b> <b>1#</b>	<b>Specimen</b> <b>2#</b>	<b>Specimen</b> <b>3#</b>	<b>Specimen</b> <b>4#</b>	<b>Specimen</b> <b>5#</b>
1	32	37	0	0	0	0	0	0
2	97	83	0	0	0	0	0	0
3	185	160	0	0	0	0	0	0
4	409	352	0	1	1	0	0	0
5	1314	1518	0	5	6	4	4	5
6	513	499	0	2	3	4	3	5
Total (T), CFU	2550	2649	<1	8	10	8	7	10
Average (C), CFU	$2.6 \times 10^3 = (P_A + P_B) / 2$							
BFE, %				99.7	99.6	99.7	99.7	99.6
Requirements	≥ 95							
Remarks	<p><i>P</i> is the value of corresponding corrected particle counts as specified by the manufacturer of the cascade impactor.  <i>T</i> is the total of <i>P</i> value for the test specimen.  <i>C</i> is the mean of the total of <i>P</i> value of the two positive controls.</p>							

## Differential pressure Test

### 1. Purpose

The purpose of the test was to measure the differential pressure of masks.

### 2. Sample description was given by client

Sample description : Single-use Medical Face Mask  
Specification : With earloop  
Lot Number : 200416  
Sample Receiving Date : 2020-05-01

### 3. Test Method

EN 14683:2019+AC:2019(E) Annex C

### 4. Apparatus and materials

Differential pressure testing instrument

### 5. Test specimen

- 5.1 Test specimen are complete masks or shall be cut from masks. Each specimen shall be able to provide 5 different circular test areas of 2.5 cm in diameter.  
5.2 Prior to testing, condition all test specimens for a minimum of 4 h at  $(21\pm 5)^{\circ}\text{C}$  and  $(85\pm 5)\%$  relative humidity.

### 6. Procedure

- 6.1 Without a specimen in place, the holder is closed and the differential manometer is zeroed. The pump is started and the flow of air adjusted to 8 L/min.  
6.2 The pretreated specimen is placed across the orifice (total area  $4.9\text{cm}^2$ , test area diameter 25mm) and clamped into place so as to minimize air leaks.  
6.3 Due to the presence of an alignment system the tested area of the specimen should be perfectly in line and across the flow of air.  
6.4 The differential pressure is read directly.  
6.5 The procedure described in steps 6.1-6.4 is carried out on 5 different areas of the mask and readings averaged.

#### Results:

Specimen	Test Results* (Pa/cm <sup>2</sup> )	Average (Pa/cm <sup>2</sup> )	Requirements	Judgement
1#	29.0	27.3	< 40	Pass
2#	27.6			
3#	24.7			
4#	25.4			
5#	29.7			



## Microbial Cleanliness Test

### 1. Purpose

The purpose of the test was to measure microbial cleanliness of mask.

### 2. Sample description was given by client

Sample description : Single-use Medical Face Mask

Specification : With earloop

Lot Number : 200416

Sample Receiving Date : 2020-05-01

### 3. Test Method

According to EN ISO 11737-1:2018 to determine the microbial cleanliness of mask material, and refer to the procedure as described in EN 14683:2019+AC:2019(E) Annex D

### 4. Apparatus and materials

4.1 Orbital shaker.

4.2 0.45 um filter.

4.3 Tryptic Soy Agar (TSA).

4.4 Sabouraud Dextrose Ager (SDA) with chloramphenicol.

4.5 Formula of Extraction Liquid: 1g/L peptone, 5g/L NaCl and 2g/L Tween 20.

4.6 Extraction apparatus.

### 5. Test specimen

5.1 As requested by client, take a total of 5 mask samples.

5.2 Mask samples for testing are provided in the original primary packaging.

5.3 Condition at (18 to 26)°C and (45 to 65)% relative humidity during testing.

### 6. Procedure

6.1 Five test specimens are selected from the top, bottom and 3 randomly chosen marks.

6.2 The mask is aseptically removed from the packaging and placed in a sterile 500 mL bottle containing 300 mL of extraction liquid.

6.3 The bottle is laid down on an orbital shaker and shaken for 5 min at 250 rpm.

6.4 After extracting, 100mL of the extraction liquid is filtered through a 0.45 um filter and laid down on a TSA plate for the total viable aerobic microbial count. Another 100 mL aliquot of the same extraction liquid is filtered in the same way and the filter plated on SDA for fungi enumeration.

6.5 The plates are incubated for 3 days at 30°C and 7 days at (20 to 25)°C for TSA and SDA plates respectively.

6.6 Calculate the colonies of each agar plate.

### 7. Calculation

For each test specimen calculate the microbial cleanliness as follows by counting the total colonies of the TSA and SDA plates.

Results\*:

Specimen	Colonies of the TSA Plate	Colonies of the SDA Plate	Microbial Cleanliness, (CFU/g)	Requirements	Judgement
1#	6	0	6	EN14683:2019+AC:2019(E) Annex D  EN ISO 11737-1:2018  ≤ 30 CFU/g	Pass
2#	6	0	6		
3#	4	0	4		
4#	3	0	3		
5#	3	0	3		

Note:

- 1.\*denotes this test was carried out by external laboratory assessed as competent.
- 2.This report is for internal use only such as internal scientific research ,education, quality control, product R&D.

-END OF THE TEST REPORT-



# Test report

Report number: 200327

Product name: Single-use medical face mask

Production lot number: 200327

Report date: 20200329

**Anhui Tiankang Medical Technology Co., Ltd.**




	After wearing the mask, it should cover the wearer's mouth, nose and jaw		
Nose clip	Masks should be fitted with nose clips made of plastic materials. Nose clip length should not be less than 8.0cm.	Meet the requirement	Pass
Mask strap	Mask straps should be adjusted easily. The breaking force at the connection point between each mask belt and the mask body should be no less than 10N.	Meet the requirement	Pass
Bacterial filtration efficiency (BFE), (%)	$\geq 95\%$	Meet the requirement	Pass
Differential pressure (Pa/cm <sup>2</sup> )	$< 40 \text{ Pa/cm}^2$	Meet the requirement	Pass
Microbial cleanliness (cfu/g)	$\leq 30 \text{ cfu / g}$	Meet the requirement	Pass
Conclusion	Comply with the provisions of EN 14683: 2019+AC: 2019 Date: March 29, 2020		

Approval:

Reviewer:

Tester:



## Kindermasken Typ I nach EN14683




## Vorderansicht





## Seitenansicht von links



## Seitenansicht von rechts





## Draufsicht



## Ansicht von hinten

